

# Richard hourigan, inc.

---

esponding to the needs of industry since 1973

## Instructions

---

### Instruction Number: 5067PH

DROP TEST  
pH (LONG RANGE)

#### COMPONENTS:

1 X 5067pH	Instruction
1 x 5425	Comparator, Color Card, pH (long range), 3.0-11.0
1 x 9017	Test Cell, Calibrated 5 mL, square, plastic w/cap
1 x R-1003U-C-DB	Long Range Indicator, 2 oz, DB

TO ORDER REPLACEMENT PARTS AND REAGENTS CALL TOLL-FREE  
877-7WATER6 (877-792-8376) or [email us](#) with your requirements.

#### PROCEDURE:

CAREFULLY READ AND FOLLOW PRECAUTIONS ON REAGENT LABELS.  
KEEP REAGENTS AWAY FROM CHILDREN.

pH (long range) (Color Comparison)

1. Rinse and fill 5 mL test cell (#9017) to 5 mL mark with water to be tested.
2. Add 5 drops R-1003U Long Range Indicator. Cap and mix.
3. Place test cell on white area of pH Long Range color card (#5425). Match color in test cell with a color standard. Record as pH units.

04/09